July 07, 2015

United States Department of the Interior



U.S. GEOLOGICAL SURVEY Reston, Virginia 20192

REPORT OF CALIBRATION

- of Acrial Mapping Camera							
Camera type: Lens type: Nominal focal Length:	Wild RC30* Wild Universal Aviogon /4-S 153 mm	Camera serial no.: Lens serial no.: Maximum aperture: Test aperture:	5364 13415 f/4 f/4				
Submitted by:	Tuck Mapping Solutions, Inc. Big Stone Gap, VA						

Reference:

These measurements were made on Agfa glass plates, 0.19 inch thick, with spectroscopic emulsion type APX Panchromatic, developed in D-19 at 68° F for 3 minutes with continuous agitation. These photographic plates were exposed on a multicollimator camera calibrator using a white light source rated at approximately 5200K.

I.	Calibrated Focal Length:	153.748 mm
----	---------------------------------	------------

II. Lens Distortion

Field angle:	7.5°	15°	22.7°	30°	35°	40°	
Symmetric radial (µm) Decentering tangential (µm)	1 0	2 0	3 0	3 1	1	-2 2	
Symmetric radial distortion	Decentering distortion				Calibrated principal point		
$\begin{array}{rcl} \kappa_{0} & = & -0.5114 \text{E-04} \\ \kappa_{1} & = & 0.1224 \text{E-08} \\ \kappa_{2} & = & 0.1788 \text{E-12} \\ \kappa_{3} & = & 0.0000 \\ \kappa_{4} & = & 0.0000 \end{array}$	P ₁ P ₂ P ₃ P ₄	= 0.1 = 0.9 = 0.0 = 0.0	487E-07 9454E-07 9000 9000		х _р Ур	= 0.002 mm = 0.007 mm	

The values and parameters for Calibrated Focal Length (CFL), Symmetric Radial Distortion $(K_0, K_1, K_2, K_3, K_4)$, Decentering Distortion (P_1, P_2, P_3, P_4) , and Calibrated Principal Point [point of symmetry] (x_p, y_p) were determined through a least-squares Simultaneous Multiframe Analytical Calibration (SMAC) adjustment. The x and y-coordinate measurements utilized in the adjustment of the above parameters have a standard deviation (σ) of ±3 microns.

^{*} Equipped with Forward Motion Compensation

III. Lens Resolving Power in cycles/mm

Area-weighted average	resolutior	109)				
Field angle:	0°	7.5°	15°	22.7°	30°	35°	40°
Radial Lines	134	159	134	134	113	95	95
Tangential Lines	134	134	113	113	113	95	80

The resolving power is obtained by photographing a series of test bars and examining the resultant image with appropriate magnification to find the spatial frequency of the finest pattern in which the bars can be counted with reasonable confidence. The series of patterns has spatial frequencies from 5 to 268 cycles/mm in a geometric series having a ratio of the 4th root of 2. Radial lines are parallel to a radius from the center of the field, and tangential lines are perpendicular to a radius.

IV. Filter Parallelism

The two surfaces of the Wild 420 filter No. 7569 and 525 filter No. 7025 accompanying this camera are within 10 seconds of being parallel. The 525 filter was used for the calibration.

V. Shutter Calibration

Indicated Time (sec)	Rise Time (μ sec)	Fall Time (µ	¹ / ₂ Width Time (ms)	Nom. Speed (sec)	Efficiency (%)
1/125	938	920	8.82	1/120	93
1/250	470	485	4.58	1/230	93
1/500	248	240	2.39	1/450	94
1/1000	135	129	1 21	1/890	93

The effective exposure times were determined with the lens at aperature f/4. The method is considered accurate within 3 percent. The technique used is described in International Standard ISO 516:1999(E).

VI. Film Platen

The platen mounted in Wild drive unit No. 5364 does not depart from a true plane by more than 13 μ m (0.0005 in).

This camera is equipped with a platen identification marker that will register "739" in the data strip area for each exposure.

USGS Report No. OSL/3672

VII. Principal Point and Fiducial Mark Coordinates



The Method of measuring these distances is considered accurate within 0.003 mm

Note: For GPS applications, the nominal entrance pupil distance from the focal plane is 277mm.

USGS Report No. OSL/3672

Stereomodel Flatness IX.

FMC Drive Unit No: 5364 Platen ID: 739





Stereomodel Test Point Array (values in micrometers)

The values shown on the diagram are the average departures from flatness (at negative scale) for two computer-simulated stereo models. The values are based on comparator measurements on Agfa Avitone P3p copy film made from Agfa Aviphot Pan 200 film exposures. These measurements are considered accurate to within 5 µm.

X. System Resolving Power on film in cycles/mm

Area-weighted average	52		Film:	Pan 200			
Field angle	0°	7.5°	15°	22.7°	30°	35°	40°
Radial Lines	57	57	57	57	57	48	48
Tangential Lines	57	57	48	57	48	48	48

This aerial mapping camera calibration report supersedes the previously issued USGS Report No. OSL/3602, dated June 27, 2012.

Ryan Longhenry Long Term Archive Project Manager Climate and Land Use Change